

Vermont Housing and Conservation Board Lead Based Paint Hazard Reduction Program

XRF Quality Control Field Check

Date:			Building Code:			
Address:					Unit:	
City			VT	Zip		
Company			Operator			
XRF Manufacturer			Model		Mode:	
Room	Wall	Component or Code	Substrate Code	Initial Reading Mg/sq.cm	QC Reading Mg/sq.cm	Difference Mg/sq.cm
Average Totals						
Average Difference Between Initial and QC Readings must be < 1.0 mg/sq. cm for average totals less than 10mg/sq.cm or ± 10% for totals > 10mg/sq.cm						

Inspected by: _____ **Certification #** _____

Is quality control adequate in this unit? **YES** **NO**